



Ref. Certif. No.

SG PSB-MD-00638

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Power supply (AC-DC converter)
Name and address of the applicant	Mornsun Guangzhou Science & Technology Co., Ltd. No. 8, Nanyun Road 4 Huangpu District 510670 Guangzhou, Guangdong PEOPLE'S REPUBLIC OF CHINA
Name and address of the manufacturer	Mornsun Guangzhou Science & Technology Co., Ltd. No. 8, Nanyun Road 4, Huangpu District, 510670 Guangzhou, Guangdong, PEOPLE'S REPUBLIC OF CHINA
Name and address of the factory	Mornsun (Huaihua) Science & Technology Co., Ltd. No. 16, Jianshe Avenue, High-tech Industrial Development District, 418000 Huaihua, Hunan, PEOPLE'S REPUBLIC OF CHINA
Ratings and principal characteristics	Rated Input: 100-240VAC, 50/60Hz, 5.2-3A or 6.5-4A Rated Output: See test report for details Protection Class: I Degree of Protection: IPX0
Model/type Ref.	See page 2
Additional information (if necessary)	See Test Report for National Differences and Group Differences. The risk management requirements of the standard were not addressed.
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1:2005/AMD2:2020
as shown in the Test Report Ref. No. which forms part of this certificate	085-220324901-000

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 105228 0033 Rev. 00

Date, 2023-01-12

(Yager Bi)

TÜV SÜD PSB Pte Ltd 15 International Business Park TÜV SÜD @ IBP Singapore 609937



PSB Singapore



Ref. Certif. No.

SG PSB-MD-00638

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Trademark / Brand (Image)



Models (continued)

LOF450-20B12, LOF450-20B12-C, LOF450-20B12-CF,
LOF450-20B12-yyy, LOF450-20B12-C-yyy, LOF450-20B12-CF-yyy,
LOF550-20B12, LOF550-20B12-C, LOF550-20B12-CF,
LOF550-20B12-yyy, LOF550-20B12-C-yyy, LOF550-20B12-CF-yyy,
LOF450-20B24, LOF450-20B24-C, LOF450-20B24-CF,
LOF450-20B24-yyy, LOF450-20B24-C-yyy, LOF450-20B24-CF-yyy,
LOF550-20B24, LOF550-20B24-C, LOF550-20B24-CF,
LOF550-20B24-yyy, LOF550-20B24-C-yyy, LOF550-20B24-CF-yyy,
LOF450-20B48, LOF450-20B48-C, LOF450-20B48-CF,
LOF450-20B48-yyy, LOF450-20B48-C-yyy, LOF450-20B48-CF-yyy,
LOF550-20B48, LOF550-20B48-C, LOF550-20B48-CF,
LOF550-20B48-yyy, LOF550-20B48-C-yyy, LOF550-20B48-CF-yyy
(y=A-Z or 0-9, stands for different marketing purpose)

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CBS 105228 0033 Rev. 00

Date, 2023-01-12

(Yager Bi)

